🕭 James Heal

EU DECLARATION of CONFORMITY

In accordance with EN ISO 17050-1:2010			
This declaration is issued under the sole responsibility of the manufacturer.			
We James Heal			
of Ri	of Richmond Works, Halifax, HX3 6EP, UK		
in accordance to the following directive(s):			
20)14/35/EU	The Low Voltage Directive	
20)14/30/EU	The Electromagnetic Compatibility Directive	
20	06/42/EC	The Machinery Directive	
hereby declare that:			
•	quipment:	Martindale	
	odels:	1600 Series: 1602 / 1605 / 1609	
	escription:	Abrasion and Pilling Tester	
	erial Number:	160X(X)/XX/XXXX	
	roduct Date:	1 st July 2016	
	oddol Dalo.		
is in conformity with the relevant Union harmonisation legislation, based on the conformity of the following documents:			
Applied Harmonised Standards: Ref. No Title Edition/date			
		Title	Edition/date
EP	N 61010-1	Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements	3 rd Ed. 2010
EN	N 61326-1	Electromagnetic Compatibility (EMC) equipment for measurement, control and laboratory use.	2013
Referenced Standards:			
Re	ef. No	Title	Edition/date
EN	N 55011	Conducted emissions (150kHz-30MHz)	2009 +A1: 2010
EN	N 61000-3-2	Harmonic emissions Class A	2006 +A1: 2009 +A2: 2009
EN	N 61000-3-3	Flicker	2008
E١	N 61000-4-2	ESD	2009
EN	N 61000-4-3	Radiated Immunity	2006 +A2: 2010
E١	N 61000-4-4	EFT/B	2004 +A1: 2010
E١	N 61000-4-5	Surge	2006
E١	N 61000-4-6	Conducted RF Immunity	2009
E١	N 61000-4-11	Voltage dips and interruptions	2004
I have by declars that the above mentioned product is in confermity with the stated Otenderde			

I hereby declare that the above mentioned product is in conformity with the stated Standards.

AUTHORISED SIGNATORY

NA Payle.

Neil Pryke

Innovation Director

DATE: 1st July 2016

The technical file is available from our EU Headquarters:

Neil Pryke Innovation Director James Heal Richmond Works Halifax, UK HX3 6EP